



**PEWA**  
Messtechnik GmbH  
Weidenweg 21  
58239 Schwerte  
Tel.: 02304-96109-0  
Fax: 02304-96109-88  
E-Mail: info@pewa.de  
Homepage : www.pewa .de

## HI 9200 Infrared Transmitter


- HI 9200/9 Infrared Transmitter with 1 m (3') cable and 9-pin serial connector
- HI 9200/25 Infrared Transmitter with 1 m (3') cable and 25-pin serial connector

### OPERATION

Using the HI 9200 Infrared Transmitter, all recorded data can be fed to your Personal Computer without having to connect and disconnect cables between your meter and computer. This ensures that there is no wear and tear on your computer. Simply attach the transmitter connector to your PC and tighten the screws. The transmitter can then be left permanently connected to your PC to facilitate wireless data transmission.


Place your data-logger on the HI 9200 Infrared Transmitter ensuring that the infrared LED's are aligned and placed right on top of each other. Press the meter down gently to

## HI 9200 Infrared Transmitter



secure a perfect contact between the infrared LED's. For downloading to your PC, follow attentively the steps specified in the instruction manual of your datalogger.

During the data transfer the meter displays "r232".



Data can be further elaborated with the new (optional) HI 92000 Windows® compatible application software.

## HI 9200 Infrared Transmitter

HI 92000 allows use of commonest spread sheet programs (e.g. Excel<sup>®</sup>, Lotus 1-2-3<sup>®</sup>) and offers a variety of features with an on-line help routine.

To install HI 92000, you need a 3.5" drive and a couple of minutes to follow the short instructions conveniently printed on the disk label.

### TROUBLESHOOTING GUIDE

If you experience difficulty in transmission, check to make sure that the transmitter is properly connected to your PC. Wipe clean the infrared LED's with a lint-free cloth on the transmitter and ensure that they are free from dust and foreign objects on the datalogger. The LED's should also be properly aligned and the datalogger should be cradled fully on the transmitter. If you need more technical information, do not hesitate to e-mail us at [tech@hannainst.com](mailto:tech@hannainst.com).

50159200 06/05

## HI 9200 Infrared Transmitter

### OPTIONAL ACCESSORIES

HI 92000 Windows<sup>®</sup> compatible application software

Windows<sup>®</sup> is a registered Trademark of "Microsoft Co."  
Excel<sup>®</sup> Copyright of "Microsoft Co."  
Lotus 1-2-3<sup>®</sup> Copyright of "Lotus Co."

Hanna Instruments reserves the right to modify the design, construction and appearance of its products without advance notice.

 **HANNA**<sup>®</sup>  
instruments  
[www.hannainst.com](http://www.hannainst.com)